



**STATE OF RHODE ISLAND
DEPARTMENT OF HEALTH
DIVISION OF EMS**

**Ambulance Minimum
Equipment List**

Effective March 1, 2008

SUMMARY BY LICENSE CLASS

The following list indicates the minimum equipment required for each class of ambulance license, along with the time permitted to correct any deficiency. A detailed definition for each item listed may be found in *Definitions*.

✓ = REQUIRED O = OPTIONAL - = NOT PERMITTED

	ITEM	A-1	A-1A	A-2	C-1	C-1A	C-2	Correction
General Requirements	Complete Tool Kit	✓	✓	✓ (1)	-	✓ (1)	-	10 days
	DOT Inspection Sticker	✓	✓	✓	✓	✓	✓	24 hours
	DOT Registration	✓	✓	✓	✓	✓	✓	24 hours
	Drive Shaft Hanger	✓	✓	✓	-	-	-	10 days
	Exhaust System	✓	✓	✓	✓	✓	✓	10 days
	Free from Rust/Dents	✓	✓	✓	✓	✓	✓	10 days
	Patient Compartment Lighting	✓	✓	✓	-	-	-	24 hours
	Non-Emergency Exterior Lighting	✓	✓	✓	✓	✓	✓	10 days
	Paint/Markings	✓	✓	✓	✓	✓	✓	10 days
	Radio Comm. with Dispatcher	✓	✓	✓	✓	✓	✓	Immediate
	Radio Comm. with Hospital	✓	✓	✓	✓	✓	-	Immediate
	Siren/ Emergency Lights	✓	✓	✓	✓	✓	✓	Immediate
	Triangle Reflectors or Flares	✓	✓	✓	O	O	O	10 days

	ITEM	A-1	A-1A	A-2	C-1	C-1A	C-2	Correction
BLS Supplies/Equipment	Adhesive Tape	✓	✓	✓	✓	✓	✓	10 days
	Blood Pressure Cuffs (all sizes)	✓	✓	✓	✓	✓	✓	Immediate
	Conforming Bandages ("Kling")	✓	✓	✓	✓	✓	✓	10 days
	Sterile Gauze (assorted sizes)	✓	✓	✓	✓	✓	✓	10 days
	Mast Trousers	✓	✓	✓	O	O	O	10 days
	Obstetrics Kit	✓	✓	✓	O	O	O	10 days
	Pediatric Dosing Device	✓	✓	✓	✓	✓	✓	10 days
	Rectal Thermometer	✓	✓	✓	O	O	O	10 days
	Trauma Shears	✓	✓	✓	✓	✓	✓	10 days
	Splints (assorted sizes)	✓	✓	✓	✓	✓	✓	10 days
	Sterile Burn Sheets (2)	✓	✓	✓	O	O	O	10 days
	Stethoscope	✓	✓	✓	✓	✓	✓	Immediate
	Traction Splint	✓	✓	✓	O	O	O	10 days
	Trauma Dressings (3)	✓	✓	✓	✓	✓	✓	10 days
	Triangular Bandages (6)	✓	✓	✓	✓	✓	✓	10 days

	ITEM	A-1	A-1A	A-2	C-1	C-1A	C-2	Correction
Extrication/Transport	Cervical Collars	✓	✓	✓	✓	✓	✓	Immediate
	Child Seat	✓	✓	✓	0	0	0	10 days
	Head Restraints	✓	✓	✓	0	0	0	10 days
	Long Spine Board	✓	✓	✓	0	0	0	Immediate
	Short Board or KED	✓	✓	✓	0	0	0	24 hours
	Straps for Boards (4)	✓	✓	✓	0	0	0	Immediate
	Stretcher/Straps/Mattress	✓	✓	✓	-	-	-	Immediate
	Triage/Tracking Tags	✓	✓	✓	0	0	0	10 days

	ITEM	A-1	A-1A	A-2	C-1	C-1A	C-2	Correction
Airway/Ventilation	Bag-Valve-Mask (all sizes)	✓	✓	✓	✓	✓	✓	Immediate
	Bite Stick	✓	✓	✓	✓	✓	✓	10 days
	Endotracheal Intubation Kit	0	0	0	0	0	0	10 days
	EOA Replacement Tube	✓	✓	✓	0	0	0	10 days
	EOA Tube w/Mask	✓	✓	✓	0	0	0	10 days
	Main Oxygen Supply	✓	✓	✓	0	0	0	Immediate
	Oxygen Cannulas	✓	✓	✓	✓	✓	✓	24 hours
	Nasopharyngeal Airways	✓	✓	✓	✓	✓	✓	Immediate
	Oropharyngeal Airways	✓	✓	✓	✓	✓	✓	Immediate
	Oxygen Masks (all sizes)	✓	✓	✓	✓	✓	✓	Immediate
	Oxygen Nebulizer	✓	✓	✓	✓	✓	✓	Immediate
	Portable Oxygen Cylinder	✓	✓	✓	✓	✓	✓	Immediate
	Portable Suction Unit	✓	✓	✓	✓	✓	✓	Immediate
	Suction Catheters w/ Tubing	✓	✓	✓	✓	✓	✓	Immediate

	ITEM	A-1	A-1A	A-2	C-1	C-1A	C-2	Correction
Infection Control	Biohazard/Refuse Container	✓	✓	✓	0	0	0	Immediate
	Body Substance Isolation Kits	✓	✓	✓	✓	✓	✓	24 hours
	Latex-Free Gloves	✓	✓	✓	✓	✓	✓	Immediate
	Level-C PPE Sets	✓	✓	✓	✓ (2)	✓ (2)	✓ (2)	10 days
	N95 Masks	✓	✓	✓	✓	✓	✓	24 hours
	Sharps Receptacle	✓	✓	✓	✓	✓	✓	Immediate

	ITEM	A-1	A-1A	A-2	C-1	C-1A	C-2	Correction
Cardiac Care	Automatic External Defibrillator	0	0	✓ (3)	0	0	✓ (3)	Immediate
	Cardiac Monitor & Defibrillator	✓	✓	0	✓	✓	0	Immediate

	ITEM	A-1	A-1A	A-2	C-1	C-1A	C-2	Correction
IV Supplies/Equipment	Blood Tubes	✓	✓	-	✓	✓	-	10 days
	Drip Sets (10-15gtts)	✓	✓	-	✓	✓	-	Immediate
	Drip Sets (60gtts)	✓	✓	-	✓	✓	-	24 hours
	IV Catheters (assorted sizes)	✓	✓	-	✓	✓	-	Immediate
	IV Infusion Pump	○	○	-	○	○	-	24 hours (6)
	IV D5W	○	○	-	○	○	-	24 hours (6)
	IV Lactated Ringer's	✓	✓	-	✓	✓	-	Immediate (4)
	IV Normal Saline	✓	✓	-	✓	✓	-	Immediate (4)
	Stopcock (3- or 4-way)	✓	✓	-	✓	✓	-	10 days
	Syringes/Needles (assorted sizes)	✓	✓	-	✓	✓	-	Immediate
	Volumetric Burette (100mL)	✓	✓	-	○	○	-	24 hours

	ITEM	A-1	A-1A	A-2	C-1	C-1A	C-2	Correction
Medications	Acetaminophen suppositories	✓	✓	✓	○	○	○	24 hours (5)
	Activated Charcoal	✓	✓	✓	○	○	○	Immediate
	Adenosine	✓	✓	-	✓	✓	-	Immediate
	Albuterol solution	✓	✓	✓	✓	✓	✓	Immediate
	Amiodarone (vials)	✓	✓	-	✓	✓	-	Immediate
	Amiodarone (admixture)	○	○	-	○	○	-	n/a
	Antacid	○	○	-	○	○	-	n/a
	Aspirin	✓	✓	✓	✓	✓	✓	Immediate
	Atropine	✓	✓	-	✓	✓	-	Immediate
	British Anti-Lewisite (BAL)	○	○	-	○	○	-	n/a
	Calcium Gluconate	○	○	○	○	○	○	n/a
	Dextrose 25%	○	○	-	○	○	-	n/a
	Dextrose 50%	✓	✓	-	✓	✓		Immediate
	Diphenhydramine (injectable)	✓	✓	-	✓	✓	-	Immediate
	Diphenhydramine (oral)	✓	✓	-	✓	✓	-	24 hours
	Epinephrine 1:10,000	✓	✓	-	✓	✓	-	Immediate
	Epinephrine 1:1000 (amps or vials)	✓	✓	✓	✓	✓	✓	Immediate
	Epi-Pen autoinjector	○	○	○	○	○	○	n/a
	Furosemide	✓	✓	-	✓	✓	-	Immediate
	Glucagon	✓	✓	○	✓	✓	○	Immediate
	Glucose (Oral)	○	○	○	○	○	○	n/a
	Hydrocortisone Sodium Succinate	✓	✓	-	✓	✓	-	Immediate
	Lidocaine (admixture)	○	○	-	○	○	-	n/a
	Lidocaine (vials)	✓	✓	-	✓	✓	-	Immediate

ITEM	A-1	A-1A	A-2	C-1	C-1A	C-2	Correction
Medications (cont'd.) Mark-I autoinjectors	✓	✓	✓	✓	✓	✓	10 days
Midazolam	✓	✓	-	✓	✓	-	Immediate
Morphine	○	○	-	○	○	-	n/a
Naloxone	✓	✓	-	✓	✓	-	Immediate
Nitroglycerin	✓	✓	-	✓	✓	-	Immediate
Sodium Bicarbonate	✓	✓	-	✓	✓	-	24 hours
Syrup of Ipecac	○	○	○	○	○	○	n/a
Terbutaline	○	○	-	○	○	-	n/a
Thiamine	✓	✓	-	✓	✓	-	24 hours
Valium (rectal gel preparation)	○	○	-	○	○	-	n/a

(1): All private ambulance companies are excluded from carrying the emergency extrication tool kit, with exception of Class A-1 or A-1A vehicles

(2): All fire apparatus are exempt from carrying Level-C PPE.

(3): A manual defibrillator is an acceptable substitute ONLY when the service is able to ensure the vehicle is staffed at all times by either ALS providers or EMT-Basics certified for Manual Defibrillation.

(4): If the vehicle has only or the other, the second has a 24-hour correction period.

(5): If the vehicle has at least half of the required amount, the remainder has a 24-hour correction period.

(6): Vehicles stocked with admixture medications must be equipped with an IV Infusion pump, IV D5W, and PVC-free administration sets.

DEFINITIONS

Where required according to *Minimum Equipment Requirements*, the following shall serve as definitions for each required item.

GENERAL REQUIREMENTS

Complete Tool Kit	Vehicle must have an emergency/extrication toolkit consisting of the following: <ul style="list-style-type: none"> 1 open-end adjustable wrench (preferably 12 inch) 1 screwdriver regular blade (preferably 12 inch) 1 screwdriver, Phillips type (preferably 12 inch) 1 hacksaw with blades 1 pliers, vise-grip type 1 hammer (3 pound) 1 crowbar (at least 24 inch) 1 battery cable cutter 2 ropes (at least ½" inch diameter, 50 feet long) 2 pair safety goggles 2 class ABC fire extinguisher at least 5 lbs. 1 pair work gloves
DOT Inspection Sticker	Vehicle must have a current DOT inspection sticker where required by Rhode Island law and DOT/RIDMV regulations.
DOT Registration	Vehicle must have current Certificate of Registration available for inspection.
Drive Shaft Hanger	Vehicle must be provided with a driveshaft hanger to protect against pole vaulting in the event of driveshaft or U-joint failure. In the case where there is a split drive shaft, a hanger must be present on both drive shafts.
Exhaust System	Vehicles exhaust system must function properly and be free of leaks and/or damage.
Free from Rust/Dents	Vehicle must be free of rust, dents and/or other body damage that impede the function of the vehicle or renders the vehicle unroadworthy.
Patient Compartment Lighting	Vehicle must have patient compartment lighting as required by applicable Federal KKK specifications. All lights must operate properly.
Non-Emergency Exterior Lighting	All exterior lighting must be operating properly including headlamps, turn signals, brake lights, reverse lights, marker lights, and clearance lights.
Paint/Markings	Vehicle must have appropriate paint and markings as required by applicable Federal KKK specifications as well as RIDOH standards.
Radio Comm. with	Vehicle must have as minimum a two-way voice communications system between

Dispatcher	the vehicle and its dispatcher.
Radio Comm. with Hospital	Vehicle must have a two-way voice communications system between the vehicle and a hospital emergency department.
Siren/ Emergency Lights	Vehicle must have a functional siren and emergency warning lights as required by applicable Federal KKK specifications and RIDOH standards.
Triangle Reflectors or Flares	Vehicle must have a minimum of three (3) triangular reflectors or road flares.

BLS SUPPLIES & EQUIPMENT

Adhesive Tape	Vehicle must be stocked with adhesive tape or equivalent in assorted sizes.
Blood Pressure Cuffs	Vehicle must be provisioned with manual blood pressure cuffs in (at a minimum) adult, child, and infant sizes. All components must be clean and in good working condition. A blood pressure set consisting of a single sphygmomanometer with multiple cuff sizes is acceptable.
Sterile Burn Sheets	Vehicle must be stocked with two or more sterile burn sheets, sealed in original, unopened manufacturer's packaging.
Conforming Bandages ("Kling")	Vehicle must be stocked with soft roll self-adhering type bandage (Kling) in assorted sizes. Must be sealed in original, unopened manufacturer's packaging.
Sterile Gauze (assorted sizes)	Vehicle must be stocked with sterile gauze pads in assorted sizes. Must be sealed in original, unopened manufacturer's packaging.
Mast Trousers	Vehicle must be stocked with at least one (1) adult size Pneumatic Anti-Shock Garment (PASG.) Non-inflatable garments are not acceptable.
Obstetrics Kit	Vehicle must be stocked with at least one (1) sterile obstetrical kit. Must be sealed in original, unopened manufacturer's packaging.
Pediatric Dosing Device	Vehicle must be stocked with at least one (1) Pediatric Dosing Device (Broselow Tape®, Pedi-Wheel® or equivalent.) Device must be 2005 AHA Compliant.
Rectal Thermometer	Vehicle must be stocked with at least one (1) rectal thermometer.
Trauma Shears	Vehicle must be stocked with at least one (1) pair of trauma shears or scissors.
Splints (assorted sizes)	Vehicle must be stocked with two (2) or more splints (may be padded board, ladder splints, or cardboard.)
Stethoscope	Vehicle must be provisioned with at least one (1) stethoscope.
Traction Splint	Vehicle must be provisioned with at least one (1) traction splinting device (i.e. Hare, Sager or equivalent.)
Trauma Dressings	Vehicle must be provisioned with at least three (3) trauma dressings. Must be sealed in original, unopened manufacturer's packaging.
Triangular Bandages	Vehicle must be provisioned with at least six (6) cloth triangular bandages.

EXTRICATION & TRANSPORT

Cervical Collars	Vehicle must be stocked with cervical collars in small, medium, large, and pediatric sizes. Adjustable collars are acceptable but must be available in both adult and pediatric sizes. Collars must be clean and in good condition with no damaged parts.
Child Seat	Vehicle must be provisioned with a child restraint seat fitting children from newborn to 4 years of age. Seat must be compliant with current Federal Motor Vehicle Safety Standards (FMVSS) requirements.
Cervical Immobilization Device (Headblocks)	Vehicle must be equipped with at least one cervical immobilization device (i.e., headblocks.) The device must be clean and free of contamination.
Long Spine Board	Vehicle must be equipped with at least one long spineboard in good condition and free of contamination.
Short Board or KED	Vehicle must be equipped with at least one (1) short spine board with all necessary accessories. Commercially prepared devices such as the Kendrick Extrication Device (KED) are acceptable.
Straps for Boards (4)	Vehicle must be equipped with at least four longboard straps in good condition and free of contamination. All straps must slide smoothly in buckles and hardware must be free of rust or other corrosion.
Stretcher/Straps/Mattress	Vehicle must be equipped with a stretcher with a mattress having a waterproof covering and 3 mounted straps.
Triage/Tracking Tags	Vehicle must be stocked with a minimum of twenty-five (25) Mettag® triage tags (or equivalent) or other RI Department of Health approved patient tracking system.

AIRWAY & VENTILATION

Bag-Valve-Masks	Vehicle must be stocked with disposable bag-valve-masks (BVMs) in adult, child, and infant sizes (at least one each.) BVMs must be clean, unused, and individually packaged.
Bite Stick	Vehicle must be provisioned with a bite stick which may be either commercial or made of tongue blades taped together and padded.
Endotracheal Intubation Kit	If equipped for oro/endotracheal intubation, kit must include laryngoscope handles and blades in adult and pediatric sizes as follows: Curved blades #2, 3, 4, 5. Straight blades #1, 2, 3, 4. Blades may be either plastic or metal but must be clean and in good condition. Handles must have fully charged batteries and a working bulb. Kit must include an assortment of uncuffed ET tubes 2.0mm to 5.0mm, cuffed ET tubes 5.5mm to 9.0mm, a flexible stylette, Magill forceps, a device to secure the ET tube in place, and an endotracheal tube placement verification device (color metric, end tidal CO ₂ , or EDD.) Tubes and stylettes must be packaged individually in unopened, sterile manufacturer's packaging.

EOA Replacement Tube	Vehicle must be stocked with a replacement EOA tube stored in clean, unopened manufacturer's packaging.
EOA Tube w/Mask	Vehicle must be stocked with a clean EOA in good condition and stored in manufacturer's packaging.
Main Oxygen Supply	Vehicle must be equipped with an onboard oxygen system having a capacity of at least 300 liters. This system must be clean, in good condition, and free of obvious leaks. The system must include at least two (2) regulators capable of delivering oxygen at flow rates ranging from 1 to 15 liters/minute.
Oxygen Cannulas	Vehicle must be stocked with nasal oxygen cannulas in adult, child, and infant sizes. Each must be sealed in original, unopened manufacturer's packaging.
Nasopharyngeal Airways	Vehicle must be stocked with an assortment of nasopharyngeal airways in sizes from 12Fr to 36Fr.
Oropharyngeal Airways	Vehicle must be stocked with at least three oropharyngeal airways in small, medium, and large sizes (size range: 40mm to 110mm.)
Oxygen Masks (all sizes)	Vehicle must be stocked with non-rebreather type oxygen masks in small, medium, and large sizes (or adult, child, and infant). Each must be sealed in original, unopened manufacturer's packaging.
Oxygen Nebulizer	Vehicle must be stocked with at least one small volume nebulizer with reservoir. Each must be sealed in original, unopened manufacturer's packaging.
Portable Oxygen Cylinder	Vehicle must be equipped with at least one portable oxygen cylinder (size D or larger) and flow regulator.
Portable Suction Unit	Vehicle must be equipped with either a hand-powered or battery-powered portable suction unit complete with disposable suction canister and all necessary tubing. If battery-powered, battery must be fully charged and unit must operate properly when tested.
Suction Catheters w/ Tubing	Vehicle must be stocked with both rigid ("tonsil-tip") and French suction catheters. French catheters must be available in a variety of sizes from 5Fr to 18Fr. All tubing and catheters must be in clean, unopened manufacturer's packaging.

INFECTION CONTROL

Biohazard/Refuse Container	Vehicle must be equipped with at least one (1) biohazard/waste receptacle with disposable liner.
Body Substance Isolation Kits	Vehicle must be stocked with at least two (2) disposable isolation kits. Each must contain (1) gown, (1) protective goggle (1) pair gloves, (1) mask, (1) head cover and (1) biohazard bag.
Latex-Free Gloves	Vehicle must be stocked with disposable latex-free gloves, preferably in multiple sizes

Level-C PPE Sets	Vehicle must be provisioned with at least two (2) Level C PPE sets consisting of a Bullard PAPR, Tychem® suit, boots, gloves, tape and carrying bag.
N 95 Masks	Vehicle must be stocked with at least five (5) N-95 Respirator Masks.
Sharps Receptacle	Vehicle must be equipped with at least one (1) rigid, disposable biohazardous sharps container. Container may not be full.

CARDIAC MONITORING & DEFIBRILLATION

Automatic External Defibrillator	<p>Vehicle must be equipped with an FDA-approved automatic external defibrillator (AED) compliant with American Heart Association standards in effect at the time of purchase. The unit must operate according to manufacturer's standards and be current on maintenance/servicing as prescribed by the manufacturer. The unit must have a fully charged battery and at least one set of pads, unopened and appropriate to device.</p> <p>* Exception: A manual defibrillator is an acceptable substitute ONLY when the service is able to ensure the vehicle is staffed at all times by either ALS providers or EMT-Basics certified for Manual Defibrillation.</p>
Cardiac Monitor & Defibrillator	<p>Vehicle must be equipped an FDA-approved cardiac monitor with manual defibrillator function and rhythm waveform display (both monophasic and biphasic types are acceptable.) The unit must be capable of synchronized cardioversion. If the unit has an AED capability, it must be compliant with American Heart Association standards in effect at the time of purchase. The unit must operate according to manufacturer's standards and be current on maintenance/servicing as prescribed by the manufacturer. The unit must have a fully charged battery. The unit must have either device-appropriate hands-free pads and/or paddles with conductive gel. The unit must have a working strip chart recorder with an adequate supply of paper. Electrodes for ECG tracings must be available in the vehicle.</p>

IV SUPPLIES & EQUIPMENT

Blood Tubes	Vehicle must be stocked with blood collection tubes in assorted sizes.
IV Drip Sets	Vehicle must be stocked with a minimum of four (4) intravenous solution administration sets (not including the 100mL Volumetric Burette defined below.) Administration sets must include at least one 60gtts set and one 10-15gtts set. All sets must in clean, unopened original manufacturer's packaging.
IV Catheters	Vehicle must be stocked with IV catheters including a minimum of two (2) each of the following: 14#, 16#, 18#, 20#, 22#, 24#. Catheters must be in clean, unopened manufacturer's original packaging.

IV Infusion Pump	Vehicle must be equipped with an IV infusion pump and PVC-free administration sets.
IV D5W	Vehicle must be stocked with at least one (1) bag of sterile intravenous solution of D5W, no smaller than 50mL. Fluid must be clear and free of precipitates. Bag must be sealed in original, unopened manufacturer's packaging.
IV Lactated Ringer's	Vehicle must be stocked with at least two (2) bags of sterile intravenous solution of Lactated Ringer's (LR), each no smaller than 250mL. Fluid must be clear and free of precipitates. Bags must be sealed in original, unopened manufacturer's packaging.
IV Normal Saline	Vehicle must be stocked with at least two (2) bags of sterile intravenous solution of Normal Saline (NS), each no smaller than 250mL. Fluid must be clear and free of precipitates. Bags must be sealed in original, unopened manufacturer's packaging.
Stopcock 3 or 4 way	Vehicle must be provisioned with at least one (1) 3-Way or 4-Way Stopcock, sealed in original, unopened manufacturer's packaging.
Syringes/Needles (assorted sizes)	Vehicle must be provisioned with syringes and needles in assorted sizes, including at least one (1) 50-60 ml syringe.
Volumetric Burette (100mL)	Vehicle must be stocked with at least one (1) volumetric burette (100mL) for IV administration. Must be sealed in original, unopened manufacturer's packaging.

MEDICATIONS

Acetaminophen suppositories	Minimum six (6) suppositories, totaling 326mg
Activated Charcoal	Minimum one (1) 50mg unit dose
Adenosine	Minimum 30 mg
Albuterol solution	Minimum one (1) 2.5mg unit dose of 0.083% solution
Amiodarone (vials)	Minimum 300mg
Amiodarone (admixture)	Minimum one (1) bag
Antacid	Optional
Aspirin	Minimum 2000mg
Atropine	Minimum 3mg
Calcium Gluconate	Optional
Cyanide antidote kit	Optional

Dextrose 25%	Optional
Dextrose 50%	Minimum 50g
Diphenhydramine (injectable)	Minimum 50mg
Diphenhydramine (oral)	Minimum 50mg
Dopamine (admixture)	Optional
Epinephrine 1:10,000	Minimum 3mg
Epinephrine 1:1000 (ampules or vial)	Minimum 5mg
Epi-Pen autoinjector	If equipped, a minimum of two (2) 0.3mg adult Epi-pen® and two (2) 0.15 Epi-pen Jr.® (May be substituted with vials, ampules, or Ana-kit®)
Furosemide	Minimum 240mg
Glucagon	Minimum one (1) unit dose
Glucose (Oral)	Optional
Hydrocortisone Sodium Succinate	Minimum two (2) 100mg doses
Lidocaine (admixture)	Optional
Lidocaine (vials)	Minimum 3mg
Mark-I autoinjectors	Minimum six (6) unit doses
Midazolam	Minimum 20mg
Morphine	Optional
Naloxone	Minimum 10mg
Nitroglycerin	Tablets or oral spray
Sodium Bicarbonate	Minimum one (1) unit dose
Syrup of Ipecac	Optional
Terbutaline	Optional
Thiamine	Minimum 100mg
Valium (rectal gel preparation)	Optional
British Anti-Lewisite	Optional